



EA MLA Signatory  
Český institut pro akreditaci, o.p.s.  
Olšanská 54/3, 130 00 Praha 3

issues

according to section 16 of Act No. 22/1997 Coll., on technical requirements for products, as amended

# CERTIFICATE OF ACCREDITATION

No. 574/2021

**KZB-Kalibrace s.r.o.**  
with registered office Mikoláše Alše 2240, 434 01 Most, Company Registration No. 03113205

to the Calibration Laboratory No. 2374  
Calibration Laboratory

Scope of accreditation:

Calibration of gauges of length, plane angle, torque, pressure, time, temperature and electrical quantities to the extent as specified in the appendix to this Certificate.

This Certificate of Accreditation is a proof of Accreditation issued on the basis of assessment of fulfillment of the accreditation criteria in accordance with

ČSN EN ISO/IEC 17025:2018

In its activities performed within the scope and for the period of validity of this Certificate, the Body is entitled to refer to this Certificate, provided that the accreditation is not suspended and the Body meets the specified accreditation requirements in accordance with the relevant regulations applicable to the activity of an accredited Conformity Assessment Body.

This Certificate of Accreditation replaces, to the full extent, Certificate No.: 683/2020 of 10. 11. 2020, or any administrative acts building upon it.

The Certificate of Accreditation is valid until: **10. 11. 2025**

Prague: 8. 11. 2021



**Lukáš Burda**  
Director of the Department  
of Testing and Calibration Laboratories  
Czech Accreditation Institute  
Public Service Company

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**KZB-Kalibrace s.r.o.**  
Calibration Laboratory  
Mikoláše Alšed 2240, 434 01 Most

**CMC for the field of measured quantity: Length**

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work-place
		min. unit	max. unit					
1	Parallel gauge blocks	0.5 mm 100 mm	to to 100 mm 500 mm		(0.8L + 0.14) µm (1L + 0.16) µm	Comparison with parallel gauge blocks	Kp 01-013	
2*	Slide gauges / slide rules, depth gauges, height gauges, gear tooth calipers	0 mm 1,000 mm	to to 1,000 mm 2,000 mm		14 µm 17 µm	Measurement of parallel gauge blocks	Kp 01-001	
3*	Linear height gauges	0 mm	to 1,000 mm		(1.2L + 0.5) µm	Measurement of parallel gauge blocks	Kp 01-001	
4*	Micrometer gauges / micrometers, pasameters, micropasameters, micrometer depth gauges	0 mm 25 mm 100 mm 1,000 mm	to to 25 mm to 100 mm to 1,000 mm		0.7 µm 1.4 µm 2.2 µm 3.8 µm	Measurement of parallel gauge blocks	Kp 01-002	
5	Two-contact and three-contact internal gauges	3 mm	to 200 mm		1.6 µm	Comparison with setting rings		
	Inside micrometer gauges	10 mm	to 1,500 mm		3.7 µm	Comparison with parallel gauge blocks		
	Micrometer gauges / inside micrometer gauges	10 mm	to 500 mm		(1L + 0.4) µm	Direct measurement by a length gauge	Kp 01-002	



Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**KZB-Kalibrace s.r.o.**  
Calibration Laboratory  
Mikoláše Alšed 2240, 434 01 Most

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work-place
		min. unit	max. unit					
	Micrometric heads	0 mm	to 100 mm		1 µm			
	Setting gauges for micrometer gauges	0 mm	to 500 mm		(1L + 0.4) µm			
		0 mm	to 950 mm		1.8 µm	Direct measurement on a linear height gauge		
6	Indicators / direct, lever indicators and internal gauges with indicator	0 mm	to 100 mm		0.5 µm	Direct measurement by a length gauge	Kp 01-003	
7*	Indicators / direct, lever indicators and internal gauges with indicator					Direct measurement on a portable measuring device	Kp 01-003	
8	Linear sensors	0 mm	to 100 mm		2.9 µm	Direct measurement by a length gauge	Kp 01-003	
9	Rules / precise gauges and measuring magnifiers	0 mm	to 100 mm		0.5 µm	Direct measurement by a length gauge	Kp 01-004	
		0 mm	to 200 mm		2.1 µm	Direct measurement by a microscope		
		200 mm	to 400 mm		2.7 µm			
		400 mm	to 600 mm		3.7 µm			
	Steel rules	0 mm	to 1,000 mm		42 µm	Comparison with a standard gauge		
		1,000 mm	to 2,000 mm		59 µm			
	Tape measures	0 m	to 8 m		0.17 mm	Comparison with a standard track		
		8 m	to 10 m		0.32 mm			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**KZB-Kalibrace s.r.o.**

Calibration Laboratory

Mikoláše Alšed 2240, 434 01 Most

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work-place
		min. unit	max. unit					
10*	Tape measures, wooden rulers and rulers of 2m length	0 m	to 100 m		(0,015L + 0,18) mm			
	Rules / steel gauges	0 mm	to 500 mm		0,12 mm	Measurement of parallel gauge blocks	Kp 01-004, Kp 01-015	
	Tape measures	0 m	to 10 m		(0,07L + 0,19) mm	Comparison with a standard gauge	Kp 01-004	
11	Setting rings and snap gauges	0,95 mm to 10 mm	to 275 mm		1 µm (4,3L + 0,7) µm	Direct and comparative measurement on a length gauge	Kp 01-005	
	Cylindrical gauges	0 mm to 100 mm	to 500 mm		0,5 µm 1 µm			
	Slot gauges	0 mm to 100 mm	to 500 mm		0,5 µm 1 µm			
	Feeler gauges and wedges	0 mm to 100 mm			0,5 µm			
	Cylindrical gauges	0 mm to 100 mm			0,5 µm			
	Measuring wires	0 mm to 10 mm			0,5 µm			
	Setting gauges for layer thickness gauges	0 mm to 25 mm			0,8 µm			
	Thread gauges – plug gauges	0 mm to 300 mm			3,1 µm			
	Thread gauges - female	5 mm to 200 mm			4 µm	Direct measurement by a microscope		
	Thread gauges - conical	0 mm to 100 mm			5 µm	Direct measurement on a length gauge and linear height gauge		

**The Appendix is an integral part of  
Certificate of Accreditation No. 574/2021 of 8/11/2021**

**Accredited entity according to ČSN EN ISO/IEC 17025:2018:**

**KZB-Kalibrace s.r.o.**

Calibration Laboratory

Mikoláše Alšed 2240, 434 01 Most

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work-place
		min. unit	max. unit					
	Efflux viscometers	0 mm	to 10 mm		4 µm	Direct measurement by a microscope		
12*	Cylindrical gauges	0 mm	to 100 mm		2 µm	Direct measurement with a micropasameter or micrometer	Kp 01-005	
	Slot gauges	0 mm	to 100 mm		2 µm			
	Feeler gauges and wedges	0 mm	to 30 mm		1.5 µm			
	Cylindrical gauges	0 mm	to 30 mm		1.5 µm			
	Thread gauges – plug gauges	0 mm	to 75 mm		5 µm			
13*	Thickness gauges and callipers	0 mm	to 500 mm	external measurement	2.2 µm	Comparison with parallel gauge blocks or thickness standard	Kp 01-010	
		3 mm	to 500 mm	internal measurement	3.7 µm	Comparison with parallel gauge blocks or setting rings		
14*	Dry layer thickness gauges	0 mm	to 25 mm	dry layers	1.4 µm	Comparative measurement by a thickness reference standard	Kp 01-009	
15	Wet layer thickness gauges	0 mm	to 15 mm	wet layers	1 µm	Direct measurement on a length gauge or a microscope	Kp 01-009	
16*	Surface plates / flatness	0 mm	to 5 mm	length up to 300 mm up to 1,000 mm up to 2,000 mm	3.7 µm 6.2 µm 34 µm	Comparison with parallel gauge blocks	Kp 01-008	
		0 mm	to 5 mm	length up to 1,000 mm up to 2,000 mm	5.1 µm 6.2 µm	Comparison with parallel gauge blocks		
		0 mm	to 5 mm	length up to 100 mm up to 300 mm	2.2 µm 2.5 µm	Comparison with parallel gauge blocks		



**The Appendix is an integral part of  
Certificate of Accreditation No. 574/2021 of 8/11/2021**

**Accredited entity according to ČSN EN ISO/IEC 17025:2018:**

**KZB-Kalibrace s.r.o.**  
Calibration Laboratory  
Mikoláše Alšed 2240, 434 01 Most

Ord. number 1	Calibrated quantity / Subject of calibration	Nominal range			Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work- place
		min.	unit	max.					
17	Roller length gauges	0 m	to	100 m	up to 500 mm up to 1,000 mm	2.8 µm 5.1 µm	Direct measurement by a special measuring device	Kp 01-014	
18	Laser distance meters	0.5 m	to	8 m		(0.2L + 10) mm 0.3 mm	Comparison with a standard measuring device	Kp 01-014	
19	Levelling rods	0 m	to	7 m		0.3 mm	Comparison with a standard track or standard tape measure	Kp 01-014	
20	Telescopic length gauges	0 m	to	7 m		0.3 mm	Comparison with a standard track	Kp 01-014	
21	Weld gauges	0 mm	to	20 mm		10 µm	Comparison with parallel gauge blocks	Kp 01-015	
22	90° angles - perpendicularity	0 mm	to	5 mm	longer side up to 100 mm up to 1,000 mm	2.8 µm (8L + 6.5) µm	Comparison with parallel gauge blocks and perpendicularity standard	Kp 02-001	
		0 mm	to	5 mm	longer side up to 100 mm up to 1,000 mm	2.2 µm 5.1 µm			
		0 mm	to	5 mm		2.9 µm			
		0 mm	to	500 mm		(1L + 0.4) µm	Direct measurement on a length gauge		
23	Measuring jigs and profile gauges	0 mm	to	500 mm		2 µm	Direct measurement on a linear height gauge	Kp 01-017	
		500 mm	to	950 mm		2 µm	Direct measurement with a micropasameter or micrometer		
		0 mm	to	100 mm		2 µm	Direct measurement with a slide gauge		
24*	Measuring jigs and profile gauges	100 mm	to	300 mm		12 µm			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**KZB-Kalibrace s.r.o.**  
Calibration Laboratory  
Mikolaše Alšed 2240, 434 01 Most

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work-place	
		min. unit	max. unit						
		300 mm	to 2,000 mm		15 µm	Comparison with parallel gauge blocks			
		2 m	to 10 m		0.5 mm	Direct measurement by a standard tape measure			
25*	Length gauges	0 mm	to 1,000 mm		(1L + 0.14) µm	Comparison with parallel gauge blocks	Kp 01-011		
26*	Measuring microscopes and profile projectors	0 mm	to 100 mm	axes X, Y	1.1 µm	Comparison with a standard gauge	Kp 01-019		
		100 mm	to 200 mm	axes X, Y	1.5 µm				
		200 mm	to 500 mm	axes X, Y	4 µm	Comparison with parallel gauge blocks			
		0 mm	to 200 mm	Z-axis	2 µm				
		0 mm	to 500 mm	perpendicularity of the X and Y axes	10 µm	Comparison with a standard square			

- 1 Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.
- 2 The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02, part of CMC, and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the value measured. If the calibration is carried out outside the laboratory premises, the measurement uncertainty may be affected.
- 3 If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).



Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**KZB-Kalibrace s.r.o.**

Calibration Laboratory

Mikoláše Alšed 2240, 434 01 Most

**CMC for the field of measured quantity: Plane angle**

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range			Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work-place
		min.	unit	max.	unit				
1*	Angle gauges	0 °		to 360 °		2,4'	Comparison with angle gauges	Kp 02-002, Kp 01-015	
2*	Measuring jigs and profile gauges	0 °		to 360 °		0,9'	Direct measurement by a microscope	Kp 01-017, Kp 01-015	
3*	Measuring microscopes and profile projectors	0 °		to 360 °		0,6'	Comparison with angle gauges	Kp 01-019	

- <sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.
- <sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02, part of CMC, and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the value measured. If the calibration is carried out outside the laboratory premises, the measurement uncertainty may be affected.
- <sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).





Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**KZB-Kalibrace s.r.o.**  
Calibration Laboratory  
Mikoláše Alšed 2240, 434 01 Most

**CMC for the field of measured quantity: Force, mechanical tests**

Ord. number 1	Calibrated quantity / Subject of calibration	Nominal range			Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work- place
		min.	unit	max.	unit				
1 *	Torque measuring devices, torque wrenches and screwdrivers, pneumatic and electric nutrunners	0,15 Nm 2 Nm 10 Nm 100 Nm		to 2 Nm to 10 Nm to 100 Nm to 1,000 Nm		Clockwise  0,67 % 0,56 % 0,46 % 0,49 %	Comparison with a torque sensor	Kp 03-001	

1 Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.  
2 The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02, part of CMC, and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %.  
3 If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).



Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**KZB-Kalibrace s.r.o.**  
Calibration Laboratory  
Mikoláše Alšed 2240, 434 01 Most

**CMC for the field of measured quantity: Pressure, mechanical stress**

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the meas. quantity		Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work-place
		min. unit	max. unit						
1*	Deformation pressure gauges, digital pressure gauges, pressure transducers and pressure measuring chains	-95 kPa	to 0 kPa	negative gauge pressure	gas	0.1 % + 61 Pa	Comparison with a pressure gauge	Kp 05-001	
		0 kPa	to 100 kPa	positive gauge pressure	gas	0.1 % + 64 Pa			
		0.1 MPa	to 0.7 MPa			0.1 % + 480 Pa			
		0.7 MPa	to 1.7 MPa			0.1 % + 1.1 kPa			
		0 MPa	to 0.7 MPa	positive gauge pressure	liquid	0.1 % + 480 Pa			
		0.7 MPa	to 1.7 MPa			0.1 % + 1.1 kPa			
		1.7 MPa	to 7 MPa			0.1 % + 6.6 kPa			
		7 MPa	to 70 MPa			0.1 % + 46 kPa			
		70 MPa	to 100 MPa			0.72 MPa			

- 1 Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.
- 2 The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02, part of CMC, and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the value measured. If the calibration is carried out outside the laboratory premises, the measurement uncertainty may be affected.
- 3 If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).



Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**KZB-Kalibrace s.r.o.**  
Calibration Laboratory  
Mikoláše Alšed 2240, 434 01 Most

**CMC for the field of measured quantity: Temperature**

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work-place
		min. unit	max. unit					
1*	Indicating thermometers and temperature measuring chains	-20 °C -5 °C 50 °C 100 °C	to to to to	-5 °C 50 °C 100 °C 650 °C	0.34 °C 0.27 °C 0.48 °C 0.64 °C	Comparison with a thermometer in a calibrating oven	Kp 07-001	
2	Non-contact thermometers	30 °C 100 °C 200 °C 300 °C 400 °C	to to to to to	100 °C 200 °C 300 °C 400 °C 500 °C	1.7 °C 2.6 °C 2.7 °C 3.0 °C 3.3 °C	Comparison with a target black body	Kp 07-002	
3*	Simulation of temperature sensor signals / temperature sensor processing units	-210 °C -100 °C 150 °C 760 °C -200 °C -100 °C 120 °C -250 °C -150 °C 0 °C 250 °C	to to to to to to to to to to to	-100 °C 150 °C 760 °C 1,200 °C -100 °C 120 °C 1,370 °C -150 °C 400 °C 250 °C 1,760 °C	J    K   T  R  K	Voltage calibrator simulation including cold junction compensation	Kp 04-001	



Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**KZB-Kalibrace s.r.o.**  
Calibration Laboratory  
Mikoláše Alšed 2240, 434 01 Most

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range			Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work-place
		min. unit	max. unit	unit					
		0 °C	to	250 °C	S	1.9 °C	Resistance calibrator simulation		
		250 °C	to	1,760 °C		1.2 °C			
		600 °C	to	1,820 °C	B	1.8 °C			
		-200 °C	to	-100 °C	N	1.0 °C			
		-100 °C	to	410 °C		0.51 °C			
		410 °C	to	1,300 °C		0.59 °C			
		-250 °C	to	-100 °C	E	1.3 °C			
		-100 °C	to	650 °C		0.47 °C			
		650 °C	to	1,000 °C		0.51 °C			
		-200 °C	to	900 °C	L	0.81 °C			
		-200 °C	to	600 °C	U	0.95 °C			
		0 °C	to	1,000 °C	C	0.72 °C			
		1,800 °C	to	2,310 °C		1.7 °C			
		-200 °C	to	0 °C		0.35 °C			
		0 °C	to	800 °C	Pt 100	0.58 °C			

- 1 Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.
- 2 The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02, part of CMC, and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the value measured. If the calibration is carried out outside the laboratory premises, the measurement uncertainty may be affected.
- 3 If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).



Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**KZB-Kalibrace s.r.o.**  
Calibration Laboratory  
Mikoláše Alšed 2240, 434 01 Most

**CMC for the field of measured quantity: Electrical quantities**

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work-place
		min. unit	max. unit					
1*	DC voltage sources	0 mV	to 100 mV		0.0037 % + 13 µV	Direct measurement by a standard multimeter	Kp 04-001	
		0.1 V	to 1 V		0.0025 % + 37 µV			
		1 V	to 10 V		0.0024 % + 0.36 mV			
		10 V	to 100 V		0.0038 % + 3.6 mV			
	DC voltage meters	100 V	to 1,000 V		0.0041 % + 36 mV	Direct generation with a standard calibrator		
		0 mV	to 100 mV		0.008 % + 12 µV			
		0.1 V	to 1 V		0.008 % + 35 µV			
		1 V	to 10 V		0.008 % + 0.35 mV			
		10 V	to 100 V		0.008 % + 3.5 mV	Direct measurement by a standard multimeter		
		100 V	to 1,000 V		0.008 % + 35 mV			
2*	Direct current sources	0 µA	to 100 µA		0.05 % + 37 nA	Direct generation with a standard calibrator		
		0.1 mA	to 1 mA		0.05 % + 0.12 µA			
		1 mA	to 10 mA		0.05 % + 2.5 µA			
		10 mA	to 100 mA		0.05 % + 14 µA			
		100 mA	to 400 mA		0.05 % + 73 µA			
		0.4 A	to 1 A		0.05 % + 0.3 mA			
		1 A	to 3 A		0.10 % + 0.8 mA			
		3 A	to 10 A		0.15 % + 1.6 mA			
		0 µA	to 100 µA		0.03 % + 35 nA	Direct generation with a standard calibrator		
		0.1 mA	to 1 mA		0.03 % + 0.12 µA			
	Direct current meters	1 mA	to 10 mA		0.03 % + 1.2 µA			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**KZB-Kalibrace s.r.o.**  
Calibration Laboratory  
Mikoláše Alšed 2240, 434 01 Most

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work-place
		min. unit	max. unit					
3*	AC voltage sources	10 mA	to 100 mA	10 Hz to 20 kHz	0.03 % + 12 $\mu$ A	Simulation using current coil		
		0.1 A	to 1 A		0.03 % + 0.18 mA			
		1 A	to 10 A		0.05 % + 2.4 mA			
		0 A	to 500 A		0.50 % + 0.49 A			
	AC voltage meters	0.1 mV	to 100 mV	10 Hz to 20 kHz	0.06 % + 54 $\mu$ V	Direct measurement by a standard multimeter	Kp 04-001	
		0.1 V	to 1 V		0.06 % + 0.36 mV			
		1 V	to 10 V		0.06 % + 3.6 mV			
		10 V	to 100 V		0.06 % + 37 mV			
		100 V	to 1,000 V		0.06 % + 0.29 V			
		0.1 mV	to 100 mV		0.08 % + 43 $\mu$ V			
4*	Alternating current sources	0.1 V	to 1 V	10 Hz to 2 kHz	0.08 % + 0.39 mV	Direct generation with a standard calibrator		
		1 V	to 10 V	10 Hz to 2 kHz	0.08 % + 4.0 mV			
		10 V	to 100 V	40 Hz to 1 kHz	0.08 % + 43 mV			
		100 V	to 1,000 V	40 Hz to 1 kHz	0.08 % + 0.60 V			
		0.1 $\mu$ A	to 100 $\mu$ A	10 Hz to 2 kHz	0.15 % + 80 nA			
		0.1 mA	to 1 mA	10 Hz to 2 kHz	0.10 % + 0.54 $\mu$ A			
		1 mA	to 10 mA	10 Hz to 2 kHz	0.15 % + 7.5 $\mu$ A			
		10 mA	to 100 mA	10 Hz to 2 kHz	0.10 % + 56 $\mu$ A			
		100 mA	to 400 mA	10 Hz to 1 kHz	0.10 % + 0.51 mA			
		0.4 A	to 1 A	10 Hz to 2 kHz	0.10 % + 0.8 mA			
	Direct measurement by a standard multimeter	1 A	to 3 A	10 Hz to 2 kHz	0.15 % + 2.4 mA			
		3 A	to 10 A	10 Hz to 2 kHz	0.15 % + 16 mA			



Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**KZB-Kalibrace s.r.o.**  
Calibration Laboratory  
Mikoláše Alšed 2240, 434 01 Most

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range			Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work-place
		min. unit	max. unit	unit					
	Alternating current meters	0.1 $\mu$ A	to	100 $\mu$ A	10 Hz to 2 kHz	0.1 % + 0.47 $\mu$ A	Direct generation with a standard calibrator		
		0.1 mA	to	1 mA	10 Hz to 2 kHz	0.1 % + 0.96 $\mu$ A			
		1 mA	to	10 mA	10 Hz to 2 kHz	0.1 % + 9.5 $\mu$ A			
		10 mA	to	100 mA	10 Hz to 2 kHz	0.1 % + 96 $\mu$ A			
		0.1 A	to	1 A	10 Hz to 2 kHz	0.1 % + 0.95 mA			
5*	DC resistance / resistors and resistance boxes	1 A	to	10 A	10 Hz to 2 kHz	0.1 % + 20 mA	Simulation using current coil	Kp 04-001 Kp 04-002	
		0 A	to	500 A	30 Hz to 60 Hz	0.34 % + 0.071 A			
		0 $\Omega$	to	10 $\Omega$		0.01 % + 12 m $\Omega$			
		10 $\Omega$	to	100 $\Omega$		0.01 % + 58 m $\Omega$			
		100 $\Omega$	to	1 k $\Omega$		0.01 % + 59 m $\Omega$			
	DC resistance meters	1 k $\Omega$	to	10 k $\Omega$		0.01 % + 0.17 $\Omega$	Direct measurement by a standard multimeter	Kp 04-001 Kp 04-002	
		10 k $\Omega$	to	100 k $\Omega$		0.01 % + 2.1 $\Omega$			
		0.1 M $\Omega$	to	1 M $\Omega$		0.01 % + 32 $\Omega$			
		1 M $\Omega$	to	10 M $\Omega$		0.04 % + 1.3 k $\Omega$			
		10 M $\Omega$	to	100 M $\Omega$		0.8 % + 35 k $\Omega$			
		100 M $\Omega$	to	1 G $\Omega$		2.0 % + 0.21 M $\Omega$			
		0.01 $\Omega$	to	0.1 $\Omega$		65 $\mu\Omega$			
		0.1 $\Omega$	to	1 $\Omega$		0.052 %			
		1 $\Omega$	to	10 $\Omega$		0.14 %			
		0 $\Omega$	to	10 $\Omega$		20 m $\Omega$			
		10 $\Omega$	to	100 $\Omega$		40 m $\Omega$	Direct generation with a standard calibrator	Kp 04-001	
		100 $\Omega$	to	1 k $\Omega$		0.27 $\Omega$			
		1 k $\Omega$	to	10 k $\Omega$		2.4 $\Omega$			
		10 k $\Omega$	to	100 k $\Omega$		24 $\Omega$			



Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**KZB-Kalibrace s.r.o.**  
Calibration Laboratory  
Mikoláše Alšed 2240, 434 01 Most

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work-place
		min. unit	max. unit					
6	Inspection equipment / insulation resistance meters	0.1 M $\Omega$	to 1 M $\Omega$	Measuring voltage up to 100 V	0.24 $\Omega$	Direct generation using a standard resistance box	Kp 04-003	
		1 M $\Omega$	to 10 M $\Omega$		5.8 k $\Omega$			
		0.1 M $\Omega$	to 0.5 M $\Omega$		0.05 % + 10 $\Omega$			
		0.6 M $\Omega$	to 0.9 M $\Omega$		0.05 % + 15 $\Omega$			
		1 M $\Omega$	to 5 M $\Omega$		0.05 % + 0.25 k $\Omega$			
		6 M $\Omega$	to 9 M $\Omega$		0.05 % + 0.35 k $\Omega$			
	meters of transition resistance	10 M $\Omega$	to 50 M $\Omega$	Measuring voltage up to 500 V	0.05 % + 2.5 k $\Omega$	Direct generation using a standard resistance box		
		60 M $\Omega$	to 100 M $\Omega$		0.05 % + 0.35 k $\Omega$			
		1 M $\Omega$	to 5 M $\Omega$		0.05 % + 0.25 k $\Omega$			
		6 M $\Omega$	to 9 M $\Omega$		0.05 % + 0.35 k $\Omega$			
		10 M $\Omega$	to 50 M $\Omega$		0.1 % + 2.5 k $\Omega$			
		60 M $\Omega$	to 100 M $\Omega$		0.1 % + 9 k $\Omega$			
	meters of transition resistance	0.1 $\Omega$	to 1 $\Omega$	Measuring voltage up to 1,000 V	0.2 % + 3.7 m $\Omega$	Direct generation using a standard resistance box		
		1 $\Omega$	to 10 $\Omega$		0.1 % + 4.7 m $\Omega$			
		10 $\Omega$	to 100 $\Omega$		0.05 % + 17 m $\Omega$			
		100 $\Omega$	to 1,000 $\Omega$		0.05 % + 0.13 $\Omega$			
		1 k $\Omega$	to 10 k $\Omega$		0.05 % + 1.3 $\Omega$			
		1 k $\Omega$	to 10 k $\Omega$		0.05 % + 1.3 $\Omega$			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**KZB-Kalibrace s.r.o.**  
Calibration Laboratory  
Mikoláše Alšed 2240, 434 01 Most

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range			Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work- place
		min. unit	max. unit						
	meters of leakage current	0.1 mA 1 mA 10 mA	to to to	1 mA 10 mA 100 mA	50 Hz to 60 Hz 50 Hz to 60 Hz 50 Hz to 60 Hz	0.10 % + 0.54 $\mu$ A 0.15 % + 7.5 $\mu$ A 0.10 % + 56 $\mu$ A	Direct measurement by a standard ammeter		

- <sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.
- <sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02, part of CMC, and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %.
- <sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).





Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**KZB-Kalibrace s.r.o.**  
Calibration Laboratory  
Mikoláše Alšed 2240, 434 01 Most

**CMC for the field of measured quantity: Time and frequency quantities**

Ord. number 1	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work- place
		min. unit	max. unit					
1*	Time interval / mechanical and digital stopwatch, timers and other time meters	5 s	to 3,600 s		11 ms	Comparison with a standard counter	Kp 06-001	

- 1 Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.
- 2 The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02, part of CMC, and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %.
- 3 If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

